

The study report was written pursuant to a Contract dated 2-10-2018 and 17-10-2018,
between the University of Siena and Light Progress

Report (version 1.0)

Siena, 26th June 2020

TEST ON UV-BOX-E3/40H-NX BY LIGHT PROGRESS



INDEX

- ❖ TARGET
- ❖ OPERATIVE SPHERE
- ❖ EQUIPMENT
- ❖ PARAMETERS ESTABLISHED FOR THE TESTS
- ❖ OPERATIVE TECHNIQUE
- ❖ DATABASE ORGANIZATION
- ❖ DATA ANALYSIS AND STATISTICS
- ❖ RESULTS
- ❖ DISCUSSION AND CONCLUSIONS
- ❖ REFERENCES
- ❖ CONTACTS
- ❖ CERTIFICATES



TARGET

To assess the effectiveness of the Light Progress UV-BOX-E3/40H-NX in inactivating selected bacterial isolates at a fixed distance and two exposure times.

OPERATIVE SPHERE

The tests were requested by Light Progress and conducted by qualified staff under the supervision of Prof. Gabriele Messina, affiliated with the Department of Molecular and Developmental Medicine, University of Siena where the tests took place.

EQUIPMENT

- Light Progress UV-BOX-E3/40H-NX (Image 1 and 2)
- 50 ml Falcon centrifuge
- Plate Count Agar
- Sterile 90 mm Ø disposable Petri dishes for bacterial cultures
- D/E neutralizing broth medium for recovery phase
- Microorganisms: *Pseudomonas aeruginosa* ATCC 27853; *Escherichia coli* ATCC 8739; *Staphylococcus aureus* ATCC 43300; *Salmonella typhimurium* ATCC 23853; *Klebsiella pneumoniae* ATCC BAA-1705
- Inoculum carriers: 20 cm² Stainless steel carriers
- Bio Class thermostat bath, Velp vortex mixer, Kartell hot plate, laminar flow hood with HEPA BIO/4 filter, KW Refrigerator +2 to +8°C, Sartorius precision balance, Nichipet EX micropipette, KW and Isco temperature chambers, Fedegari sterilising autoclave, Sigma phosphate buffered saline (PBS), sterile polypropylene tubes, sterile spatula, sterile pliers, sterile glass bottles, various glassware, centrifuge
- Microsoft Excel 2016 for data collection
- Stata SE/16.0 for statistical analysis





Image 2 Light Progress UV-BOX-E3/40H-NX with light OFF



Image 1 Light Progress UV-BOX-E3/40H-NX with light ON

REPORT V 1.0 "TEST ON UV-BOX-E3/40H-NX BY LIGHT PROGRESS"



PARAMETERS ESTABLISHED FOR THE TESTS

Exposure time: i) in the preliminary stage we tested several exposure times; ii) the main tests were conducted at 2 and 3 minutes

Distance from source: 33 cm (from upper light source)

Repetitions: testing was performed three times in triplicate between May and June 2020

Concentrations: 1.5×10^7 ; 1.5×10^6 CFU/mL

OPERATIVE TECHNIQUE

Propagation of the microorganisms was conducted according to standard operating procedure (ANNEX 1 for details, separate document). The inoculum was re-suspended in test culture at concentrations of approx. 1.5×10^6 and 1.5×10^7 CFU/mL.

Set Up

Stainless steel carriers were inoculated with 100 µl of test culture. The inoculum was spread with a sterile spatula over approximately 20 cm² of each carrier and left to dry inside the laminar flow hood.

Carriers were placed as follows:

- 1) Treated sample: 2 horizontal supports (1 per bacterial concentration) were placed in the UV-box, 33 cm from the upper light source of the device.
- 2) Positive control: 2 supports (1 per bacterial concentration) were left in the lab, out of range of UV radiation.

At the end of exposure, exposed and non-exposed carriers were transferred to 90 mm Petri dishes and 10 mL D/E medium was added. Subsequently, the D/E medium was transferred to a 50 mL Falcon centrifuge and spun for 40 minutes at 4500 rpm. The supernatant was eliminated and the pellet re-suspended in 1 mL D/E medium. Finally, 100 µl was transferred to Plate Count Agar and incubated at 36°C for 48 h.



DATABASE ORGANIZATION

The first tests were conducted in an exploratory way so as to standardize the operative protocol and exposure times. Exposures between 1 and 20 minutes were tested for all microbes (ANNEX 1 for details, separate document). The following variables collected during the study were entered in a database:

- Petri dish ID
- CFUs/mL
- Microorganism species
- Inoculum concentrations

DATA ANALYSIS AND STATISTICS

Data analysis and statistical computations were supervised by Prof. Gabriele Cevenini, Department of Medical Biotechnologies, Laboratory of Applied Bioengineering – IT Engineering in Medicine, University of Siena. Microsoft Excel software (ver. 16) was used for preliminary statistical evaluation of empirical data and to organize the database. Analysis was conducted using Stata software Ver 16. The results of each experiment in triplicate were expressed as mean CFU/mL for each test. The mean logarithmic reduction and the 95% confidence interval were evaluated from the replicate data of each microbe. The tables created show the mean logarithmic reduction in CFU/mL of each microorganism compared to the real positive controls (Tables A) and theoretical positive controls (Tables B), as maximum number of CFU/mL (1.5×10^7 CFU/mL and 1.5×10^6 CFU/mL).



RESULTS

The results for the different microorganisms are shown in Tables 1A to 5B.

Table 1A. *Staphylococcus aureus* ATCC 43300 CFU/mL logarithmic reduction (real positive control reference)

2 minutes				3 minutes				
1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	mean	95% IC	mean	95% IC	
Log_{10} reduction	6.68	5.62 - 7.74	6.22	6.04 - 6.39	7.22	7.04 - 7.39	6.22	6.04 - 6.39

Table 1B. *Staphylococcus aureus* ATCC 43300 CFU/mL logarithmic reduction (theoretical positive control reference)

2 minutes				3 minutes				
1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	mean	95% IC	mean	95% IC	
Log_{10} reduction	6.94	5.90 - 7.99	6.48	6.48 - 6.48	7.04	6.19 - 7.89	6.48	6.19 - 6.48



Table 2A. *Escherichia coli* ATCC 8739 CFU/mL logarithmic reduction (real positive control reference)

2 minutes		3 minutes	
1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)
mean	95% IC	mean	95% IC
Log ₁₀ reduction	7.45	7.30 - 7.59	6.45
		6.30 - 6.59	7.45
			3.95 - 7.59
			6.45
			6.30 - 6.59

Table 2B. *Escherichia coli* ATCC 8739 CFU/mL logarithmic reduction (theoretical positive control reference)

2 minutes		3 minutes	
1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)
mean	95% IC	mean	95% IC
Log ₁₀ reduction	7.48	7.48 - 7.48	6.48
		6.48 - 6.48	7.48
			7.48 - 7.48
			6.48
			6.48 - 6.48





Table 3A. *Pseudomonas aeruginosa* ATCC 27853 CFU/mL logarithmic reduction (real positive control reference)

	2 minutes			3 minutes				
	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)		
mean	95% IC	mean	95% IC	mean	95% IC	mean	95% IC	
Log_{10} reduction	7.05	6.87 - 7.22	6.05	5.87 - 6.22	7.05	6.87 - 7.22	6.05	5.87 - 6.22

Table 3B. *Pseudomonas aeruginosa* ATCC 27853 CFU/mL logarithmic reduction (theoretical positive control reference)

	2 minutes			3 minutes				
	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)		
mean	95% IC	mean	95% IC	mean	95% IC	mean	95% IC	
Log_{10} reduction	7.48	7.48 - 7.48	6.48	6.48 - 6.48	7.48	7.48 - 7.48	6.48	6.48 - 6.48





Table 4A. *Salmonella typhimurium* ATCC 23853 CFU/mL logarithmic reduction (real positive control reference)

	2 minutes			3 minutes				
	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	mean	95% IC		
mean	95% IC	mean	95% IC	mean	95% IC	mean		
Log₁₀ reduction	6.03	4.74 - 7.32	5.03	3.74 - 6.32	6.46	4.95 - 7.97	5.66	4.61 - 6.71

Table 4B. *Salmonella typhimurium* ATCC 23853 CFU/mL logarithmic reduction (theoretical positive control reference)

	2 minutes			3 minutes				
	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	mean	95% IC		
mean	95% IC	mean	95% IC	mean	95% IC	mean		
Log₁₀ reduction	6.31	5.05 - 7.56	5.31	4.05 - 6.56	6.74	5.30 - 8.18	5.94	4.90 - 6.99

Table 5A. *Klebsiella pneumoniae* ATCC BAA-1705 CFU/mL logarithmic reduction (real positive control reference)

	2 minutes			3 minutes			
	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^6 (CFU/mL)	
mean	95% IC	mean	95% IC	mean	95% IC	mean	
Log_{10} reduction	6.48 - 5.85 - 7.11	6.01	5.50 - 6.52	7.01	6.50 - 7.52	6.01	5.50 - 6.52

Table 5B. *Klebsiella pneumoniae* ATCC BAA-1705 CFU/mL logarithmic reduction (theoretical positive control reference)

	2 minutes			3 minutes			
	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^7 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^6 (CFU/mL)	1.5×10^6 (CFU/mL)	
mean	95% IC	mean	95% IC	mean	95% IC	mean	
Log_{10} reduction	6.94 - 5.90 - 7.99	6.48	6.48 - 6.48	7.48	7.48 - 7.48	6.48	6.48 - 6.48



DISCUSSION AND CONCLUSIONS

The inoculum spread on the steel plates was coherent with those described in the literature and proposed by the manufacturer.

The Light Progress UV-BOX-E3/40H-NX-R had a biocidal effect on all microbes at exposures of 2 and 3 minutes.

The higher bacterial inactivation effect, $6\text{-}7 \log_{10}$, was achieved with 3 minutes of exposure for all five strains.

Different concentrations and dilutions on the carriers could produce different results (see ANNEX 1 for details, separate document).

Different distances and exposure times could produce different results (see ANNEX 1 for details, separate document).

Further tests using different exposure times are underway in order to implement the current CFU/mL logarithmic reduction model for all five strains.



REFERENCES

- ALTMAN, D.G., 1990: *PRACTICAL STATISTICS FOR MEDICAL RESEARCH*, CHAPMAN AND HALL/CRC.
- BOYCE J.M., DONSKEY C.J. UNDERSTANDING ULTRAVIOLET LIGHT SURFACE DECONTAMINATION IN HOSPITAL ROOMS: A PRIMER; INFECT CONTROL HOSP EPIDEMIOL. 2019 SEP;40 (9):1030-1035. DOI: 10.1017/ICE.2019.161. EPUB 2019 JUN 18.
- EVERITT, B. AND PALMER, C.R., 2006: *ENCYCLOPAEDIC COMPANION TO MEDICAL STATISTICS*, HODDER ARNOLD.
- KOWALSKI, W., 2009: ULTRAVIOLET GERMICIDAL IRRADIATION HANDBOOK: UVG FOR AIR AND SURFACE DISINFECTION, SPRINGER.
- NICOLETTI, G. AND NICOLOSI, V.M., 1998: *DIZIONARIO DI BATTERIOLOGIA UMANA NORMALE E PATOLOGICA*, MOMENTO MEDICO, MILAN.
- ZIMBRO, M.J. ET AL., 2009: *DISCO & BBL MANUAL –MANUAL OF MICROBIOLOGICAL CULTURE MEDIA-* SECOND EDITION.

CONTACTS

Prof. Gabriele MESSINA, University of Siena, Department of Molecular and Developmental Medicine, Via A. Moro 2, 53100 Siena.
Phone: +39-(0)577-235-423; Fax: +39-(0)577-234-090; Mobile: + 39-339-6699-422; Email: gabriele.messina@unisi.it

Prof. Gabriele Messina






UNIVERSITÀ
DI SIENA
1240

Department of Molecular and Developmental Medicine



CERTIFICATO



per il sistema di gestione secondo
EN ISO 9001:2015

La comprova dell'applicazione conforme ai criteri normativi è stata
consegnata e viene attestata secondo la procedura TÜV AUSTRIA CERT per

**Università degli Studi di Siena
Dipartimento di Medicina Molecolare e dello Sviluppo
EpidMol**

**Sede legale:
IT-53100 Siena (SI), Via Banchi di Sotto, 55**

**Sede operativa:
IT-53100 Siena (SI), Via Aldo Moro, 2**

Campo di applicazione

**Titolazioni di anticorpi mediante analisi sierologiche;
diagnosi microbiologiche; analisi epidemiologiche;
determinazioni chimico/biologiche ambientali.**

N° registrazione certificato: 20100193005061

Valido fino al 2022-02-05

Prima certificazione: 2019-02-06

Organismo di Certificazione
del TÜV AUSTRIA CERT GMBH

Vienna, 2019-02-06

Questa certificazione è stata eseguita secondo la procedura TÜV AUSTRIA CERT per verifiche e
certificazioni e viene periodicamente sorvegliata.

TÜV AUSTRIA CERT GMBH Deutschstraße 10 A-1230 Wien www.tuv.at



000391-18-4

REPORT V 1.0 "TEST ON UV-BOX-E3/40H-NX BY LIGHT PROGRESS"



Verifying this seal online: www.tuv-austria.com | The reproduction of this document is subject to the agreement by TÜV AUSTRIA





UNIVERSITÀ
DI SIENA
1240

ANNEX 1:
**TEST ON UV-BOX-E3/40H-NX
BY LIGHT PROGRESS**



MICROBES

- *Pseudomonas aeruginosa* ATCC 27853
- *Escherichia coli* ATCC 8739
- *Stafilococcus aureus* ATCC 43300
- *Salmonella Typhimurium* ATCC 23853
- *Klebsiella pneumoniae* ATCC BAA-1705




Operative Protocol

Light Progress Experiments: UV-BOX-E3/40H-NX-R

- Prepare a 0.5 McFarland inoculum of the bacterial strain (use PBS solution to suspend the colonies).
- Prepare two dilution of the inoculum: 1×10^{-8} and 1×10^{-7} (use PBS solution for the dilutions).
- Spread 100 μl of each dilutions on the stainless steel carriers.
- Place 1 supports for each strain dilutions inside the UV device (TREATED SAMPLES).
- Place 1 more supports (1 per strain dilutions) out of the device range (POSITIVE CONTROLS).
- Set the times in minutes and start the exposure.
- At the end of the exposure, transfer both the exposed and non-exposed supports into 90 mm Petri dishes and add 10 ml of D/E medium.
- After initial shake, let the carrier inside the petri dishes with the medium for 10 minutes.
- Transfer the D/E medium to a 50 ml Falcon and centrifuge for 40 minutes at 4500 rpm.
- Eliminate the supernatant and suspend the pellet in 1 ml of initial D/E medium.
- Transfer 100 μl to a Plate Count Agar and spread until dry (colony count have to be multiplied by 10 considering that we take 1/10 of 1ml).
- Incubate the plates at 36°C for 48h.



SCHEMATIC PROCEDURE





the following photos represent an example of petri dishes from each experiment



Staphylococcus aureus ATCC 43300: First test

- Concentrations:**
- 1×10^7 CFU/mL
- 1×10^6 CFU/mL
- Times:**
- 15, 20 minutes
- Distance:**
- 33 cm (from upper light source)
- Medium:**
- PCA (Plate Count Agar)
- Incubation:**
- 36 °C for 48h



SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

Preliminary stage Tests

Stafilococcus aureus ATCC 43300:
Second test

Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 5, 10, 15 minutes

Distance:

- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:


- 36 °C for 48h



SMP: Sample
CTR: Positive Control
Dil: Dilution
NC: Not Countable

Preliminary stage Tests

Staphylococcus aureus ATCC 43300:
Third test




SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

UNIVERSITÀ
DI SIENA
1240

Preliminary and part of main
stage Tests



Escherichia coli ATCC 8739:
First test



Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 5, 10, 15, 20 minutes

Distance:


- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h



Preliminary stage Tests

SMP: Sample
 CTR: Positive Control
 DIL: Dilution
 NC: Not Countable

Escherichia coli ATCC 8739:
Second test

Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 5, 10 minutes

Distance:



-33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h



SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

Preliminary stage Tests

Escherichia coli ATCC 8739:
Third test



Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1,2,3 minutes

Distance:


- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h



SMP: Sample

CTR: Positive Control


DIL: Dilution

NC: Not Countable



Pseudomonas aeruginosa ATCC 27853:

First test



Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 5, 10, 15, 20 minutes

Distance:


- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h



SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

Preliminary stage Tests

Pseudomonas aeruginosa ATCC 27853:
Second test



Concentrations:

- Times:**

 - 1x10⁷ CFU/mL
 - 1x10⁶ CFU/mL
 - 2,3,5 minutes

Distance:

- 33 cm (from upper light source)


Medium:

- PCA (Plate Count Agar)

Inception:

- 36 °C for 48 h

- 36 °C for 48h



SMP: Sample

CTR: Positive Con-

DILUTION

NC: Not Countable

UNIVERSITY LIBRARIES

UNIVERSITÀ
DI SIENA 1740

Preliminary and part of main stage Tests

Pseudomonas aeruginosa ATCC 27853:
Third test

Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1,2,3 minutes

Distance:



- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h



SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable



Salmonella typhimurium ATCC 23853:
First test

Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 5, 10, 15 minutes



Distance:
- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h



SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

Preliminary stage Tests

Salmonella typhimurium ATCC 23853:

Second test

Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1,2,3 minutes


Distance:
- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36°C for 48h



SMP: Sample

CTR: Positive Control

DIL: Dilution

NC: Not Countable

Salmonella typhimurium ATCC 23853:
Third test



Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1,2,3 minutes

Distance:

- 33 cm (from upper light source)

Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h




SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

Preliminary and part of main
stage Tests
UNIVERSITÀ
DI SIENA
1240



Klebsiella pneumoniae ATCC BAA-1705:
First test



Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1, 2, 3 minutes


Distance:
- 33 cm (from upper light source)

Medium:

- CLED Agar (Cystine-Lactose-Electrolyte-Deficient)

Incubation:

- 36 °C for 48h



SMP: Sample

CTR: Positive Control

DIL: Dilution

NC: Not Countable

UNIVERSITÀ
DI SIENA
1240

Preliminary and part of main

stage Tests



Klebsiella pneumoniae ATCC BAA-1705:

Second test

Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1,2,3 minutes


Distance:
- 33 cm (from upper light source)

Medium:


- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h




SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable



Preliminary and part of main
stage Tests

Klebsiella pneumoniae ATCC BAA-1705:
Third test



Concentrations:

- 1×10^7 CFU/mL
- 1×10^6 CFU/mL

Times:

- 1, 2, 3 minutes

Distance:

- 33 cm (from upper light source)


Medium:

- PCA (Plate Count Agar)

Incubation:

- 36 °C for 48h

Possible external contamination



SMP: Sample
CTR: Positive Control
DIL: Dilution
NC: Not Countable

Preliminary and part of main
stage Tests